

RapID™ NF PLUS QUALITY CONTROL REPORT FORM

1. *Record results of biochemical reactions under the expected reactions.
 Corrective Action: See Reverse Page

Date QC performed: _____
 Technologist: _____
 NF Plus lot #: _____
 Expiration Date: _____
 Approved by: _____

Organism	ADH	TRD	EST	PHS	NAG	aGLU	BGLU	ONPG	URE	GLU	PRO	PYR	GGT	TRY	BANA	IND	NO3
A. baumannii	-	-	+	(-)	-	-	-	-	-	(+)	-	-	-	(-)	-	-	-
ATCC® 19606																	
	ADH	TRD	EST	PHS	NAG	aGLU	BGLU	ONPG	URE	GLU	PRO	PYR	GGT	TRY	BANA	IND	NO3
A. hydrophilia ^a	+	+	+	+	+	-	+	+	(-)	+	+	(-)	(-)	-	(-)	+	+
ATCC® 35654																	
	ADH	TRD	EST	PHS	NAG	aGLU	BGLU	ONPG	URE	GLU	PRO	PYR	GGT	TRY	BANA	IND	NO3
E. meningoseptica	(-)	(-)	V	(+)	+	+	(+)	(-)	-	-	+	+	+	+	+	+	-
ATCC® 13253																	
	ADH	TRD	EST	PHS	NAG	aGLU	BGLU	ONPG	URE	GLU	PRO	PYR	GGT	TRY	BANA	IND	NO3
O. ureolytica ^a	(-)	-	-	-	-	-	-	-	+	-	V	-	+	(-)	-	-	(+)
ATCC® 43534																	

+ = Positive V = Variable
 - = Negative (-) = Usually negative
 (+) = Usually positive

*Visual Interpretation Only

^aKey indicator strain

CORRECTIVE ACTION

1. List organism with biochemical reaction that is outside the acceptable criteria:

2. Date corrective action started: _____

3. Corrective action procedure performed: _____

4. Results after corrective action:

Acceptable Reaction

Outside Acceptable Reaction

5. Further corrective action results, if the biochemical reaction is still outside the acceptable criteria:

6. Date corrective results accepted: _____